

APPENDIX A
NOTICE OF VIOLATION

Veterans Administration Medical Center
Castle Point, New York 12511

Docket No. 030-02626
License No. 31-11399-02

As a result of the inspection conducted on February 7, 1985, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violation was identified:

License Condition No. 15 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in application received November 8, 1979.

Item 10 of this application requires that you calibrate your survey instruments in accordance with the procedures in Appendix D, Section 1, of Regulatory Guide 10.8, Revision 1.

Item A.3 of Appendix D, Section 1 requires that survey meters be calibrated at least annually and after servicing.

Contrary to this requirement, as of the date of this inspection, the survey meter had not been calibrated since January, 1982, a period in excess of one year.

This is a Severity Level IV violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Veterans Administration Medical Center is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

8505310351 850524
REQ1 LIC30
31-11399-02 PDR