

FEB 11 1993

Advanced Medical Systems, Inc.
ATTN: Sherry Stein, Director
of Regulatory Affairs
121 N. Eagle Street
Geneva, OH 44041

License No. 34-19089-01
Docket No. 030-16055

Dear Ms. Stein:

This refers to the routine safety inspection conducted by Mr. Wayne Slawinski of this office on January 19-20, 1993, of activities authorized by NRC Byproduct Material License No. 34-19089-01, and to the discussion of our findings with Sherry Stein and Edward Svigel at the conclusion of the inspection.

The inspection was an examination of certain activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and with the conditions of your license. The inspection was conducted primarily to review your teletherapy service activities and consisted of a selective examination of procedures and representative records, and interviews with personnel.

No violations of NRC requirements were identified during the course of this inspection. However, we are concerned that some recent teletherapy service reports compiled by your service engineer were incomplete, lacking such information as survey meter and teletherapy head identification information and source manufacturer's name and model number. This is of particular concern since similar problems were identified in February 1992, during an audit of the service program by your consultant. Although service reports for 1992 were improved over those from 1991, more attention to these reports is necessary. In addition, we are concerned that your procedure 13P-21 for notifying Agreement States of service activities, does not contain current Agreement State information. Consequently, please describe your actions to address these two concerns. Your written response is requested within 30 days from the date of this letter.

In accordance with 10 CFR 2.790 of the Commission's regulations, a copy of this letter and your response to this letter will be placed in the NRC Public Document Room.

9302220174 930211
PDR ADOCK 03016055
C PDR

IEO 7
110

FEB 11 1993

The response directed by this letter is not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, PL 96-511.

We will gladly discuss any questions you have concerning this inspection.

Sincerely,

Original Signed by Roy J. Caniano

Roy J. Caniano, Chief
Nuclear Materials Safety
Branch

cc: DCD/DCB (RIDS)
Mark Loeser, RSO
1020 London Rd.
Cleveland, Ohio

(yes)
RIII

Stawinski/jaw

2/14/93

RIII

Holt

2/14/93

RIII

Caniano