

MAY 28 1985

Docket Nos. 30-00494  
30-12571

License Nos. 37-13187-01  
37-13187-02

Grandview Hospital  
ATTN: Ms. Dorothy C. Gehres  
Assistant Administrator  
700 Lawn Avenue  
Sellersville, Pennsylvania 18960

Gentlemen:

Subject: Inspection Report Nos. 030-00494/85-01 and 030-12571/85-01

This refers to the routine safety inspection conducted by Mrs. Teresa Hall Darden of this office on February 6, 1985, of activities authorized by NRC License Nos. 37-13187-01 and 37-13187-02 and to the discussions of our findings held by Mrs. Darden with yourself at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of your licenses. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, measurements made by the inspector, and observations by the inspector.

Our inspector also verified the steps you have taken to correct the violations brought to your attention in the enclosure to our letter dated March 10, 1983. We have no further questions regarding your action at this time.

Based on the results of this inspection, it appears that certain of your activities were not conducted in full compliance with NRC requirements, as set forth in the Notice of Violation, enclosed herewith as Appendix A. These violations have been categorized by severity level in accordance with the revised NRC Enforcement Policy (10 CFR 2, Appendix C) published in the Federal Register Notice (49 FR 8583) dated March 8, 1984. You are required to respond to this letter and in preparing your response, you should follow the instructions in Appendix A.

From the discussions conducted during the inspection, we understand that your area survey procedures will be upgraded to include daily surveys of the Nuclear Medicine laboratory. Further we understand that the individuals who work in the Nuclear Medicine Department will monitor themselves for contamination after each procedure or prior to leaving the Department and maintain a record of the results. We will review your actions on these matters during the next inspection of your licensed activities.

8506030194 850528  
REG1 LIC30  
37-13187-01 PDR

OFFICIAL RECORD COPY

IR GRANDVIEW 85-01 - 0001.0.0  
05/20/85

RETURN ORIGINAL TO  
REGION I

IE:07

MAY 28 1985

Grandview Hospital

2

In accordance with Section 2.790 of the NRC's "Rules of Practice," Part 2, Title 10, Code of Federal Regulations, a copy of this letter and your reply will be placed in the Public Document Room.

The responses directed by this letter and the accompanying Notice are not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, PL 96-511.

Your cooperation with us in this matter is appreciated.

Sincerely,

Original Signed By:  
John D. Kinneman

John D. Kinneman, Chief  
Nuclear Materials Safety Section A,  
Division of Radiation Safety  
and Safeguards

Enclosure:  
Appendix A, Notice of Violation

cc w/encl:  
Public Document Room (PDR)  
Nuclear Safety Information Center (NSIC)  
Commonwealth of Pennsylvania

bcc w/encl:  
Region I Docket Room (w/concurrences)  
Senior Operations Officer (w/o encl)  
T.H. Darden

RI:DRSS  
Darden/sm  
05/23/85

RI:DRSS  
Kinneman  
05/23/85

OFFICIAL RECORD COPY

IR GRANDVIEW 85-01 - 0002.0.0  
05/20/85