

APPENDIX A

NOTICE OF VIOLATION

Grandview Hospital  
Sellersville, Pennsylvania

Docket No. 030-00494  
License No. 37-13187-01

As a result of the inspection conducted on February 6, 1985, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

- A. 10 CFR 35.22(c) requires that monthly spot checks on teletherapy units not performed by a qualified expert be reviewed by a qualified expert within 15 days after performance of the checks.

Contrary to the above, as of February 6, 1985, records were not available to verify that monthly spot-checks had been reviewed by a qualified expert since October, 1983.

This is a Severity Level IV violation. (Supplement VI)

- B. 10 CFR 35.23(b) requires that the dosimetry system used to perform the spot-check measurements on a teletherapy unit required by §35.22(b)(4) be calibrated in accordance with §35.23(a) or by direct intercomparison with a system that has been calibrated in accordance with §35.23(a).

Contrary to the above, as of February 6, 1985, records were not available to verify that the pocket dosimeter used to perform the spot-check measurements required by §35.22(b)(4) had been intercompared with a system calibrated in accordance with 10 CFR 35.23.

This is a Severity Level IV violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Grandview Hospital is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

B506030199 B50528  
REG1 LIC30  
37-13187-01 PDR