

APPENDIX A

NOTICE OF VIOLATION

Nebraska Public Power District
Cooper Nuclear Station

Docket: 50-298/85-14
License: DPR-46

Based on the results of an NRC inspection conducted during the period of April 1-5, 1985, and in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C), 49 FR 8583, dated March 8, 1984, the following violation was identified:

Radiation Control Procedures

Technical Specification 6.3.4 requires procedures to be maintained consistent with the requirements of 10 CFR Part 20. Station Procedure 9.3.1.2.2, Section VI.B.10 and 11, contains requirements for operation and calibration of extender probe survey meters. This procedure requires that each scale on the extender probe-extender Model 1000 W, be calibrated at three points.

Contrary to the above, during the period February 1984 through November 1984, the licensee failed to perform calibrations at three points on each scale of extender probe-extender Model 1000W, Serial Numbers 15684, 15702, 15704, 15705, 15706, 15709, 15918, and 15920 radiation monitoring instruments.

This is a Severity Level V Violation. (Supplement IV) (50-298/8514-01)

Pursuant to the provisions of 10 CFR 2.201, Nebraska Public Power District is hereby required to submit to this office, within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated: MAY 29 1985

8506030379 850529
PDR ADOCK 05000298
G PDR