



Divine Providence Hospital 1100 Grampian Blvd., Williamsport, Pa. 17701 (717) 326-8181

May 8, 1985

Mr. John Kinneman
Nuclear Materials Safety Section A
Division of Radiation Safety and
Safeguards
Nuclear Regulatory Commission
Region I
631 Park Avenue
King of Prussia, PA 19406

Re: License No's. 37-16101-01 and
37-16101-02

Dear Mr. Kinneman:

We are in receipt of your letter dated April 25, 1985, along with the attached notice. The violation is being corrected as follows:

The Keithley dosimetry system has been given to the National Bureau of Standards (on March 13, 1985) for calibration. It is expected within a few weeks. Upon receipt, we should be in full compliance. A second system is being purchased in light of the backlog at NBS. This step should help us avoid this from happening again.

Thank you for your assistance.

Sincerely,

Richard C. Gibson
Assistant Administrator

RCG/mm

Attach.

8506030397 850528
REG1 LIC30
37-16101-01 PDR

RECEIVED-REGION I
MAY 13 PM 3:41