

APPENDIX A

NOTICE OF VIOLATION

John B. Pierce Foundation of Connecticut, Inc.
New Haven, Connecticut 06519

Docket No. 030-14580
License No. 06-02354-02

As a result of the inspection conducted on April 9, 1985, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violation was identified:

Condition 16 of License No. 06-02354-02 requires that licensed material be possessed and used in accordance with statements, representations and procedures contained in an application dated May 10, 1983, and the letters dated August 16, 1983, and November 7, 1983.

Item 1.d. of the letter dated August 16, 1983 requires that survey meters be calibrated every six months.

Contrary to the above, as of April 9, 1985, a Victoreen 470 survey meter had not been calibrated since September 28, 1981.

This is a Severity Level IV violation. (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, John B. Pierce Foundation of Connecticut, Inc. is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

8506030398 850528
REG1 LIC30
06-02354-02 PDR