



COLE HOSPITAL

809 WEST CHURCH STREET • CHAMPAIGN, ILLINOIS 61820 • 217/351-7200

May 9, 1985

Bruce S. Mallett, Ph.D.
Materials Licensing Section
U.S. Nuclear Regulatory Commission
799 Roosevelt Road
Glen Ellyn, IL 60137

RE: Request to amend NRC License Number 12-20467-01
for Termination

Dear Dr. Mallett:

In response to your letter dated February 21, 1985, we have requested the personnel of Nuclear Diagnostics, Inc. and Radiation Protection Consultants, Ltd., to perform the close-out surveys for our license termination. These individuals have informed us that the close-out surveys will be performed on or about the last two weeks of May, 1985.


Upon the completion of the close-out survey, Nuclear Diagnostics will submit the following:

1. Description of surveys.
2. Description of instruments used and methods used to convert wipe test results to disintegrations per minute per unit area.
3. Results of all tests.

We will forward this information in duplicate to your attention.

If you have any questions, please contact the undersigned. Thank you for your cooperation in this matter.

Very truly yours,


Ted McLendon
Acting Administrator

/mb
cc: Nuclear Diagnostics, Inc.

RECEIVED
JUN 28 1985
REGION III

8507260301 850712
REG3 LIC30
12-20467-01 PDR

JUN 28 1985



RADIATION PROTECTION CONSULTANTS, LTD.

Serving Radiology, Nuclear Medicine, Therapy and Industrial Facilities Since 1968

710 EAST OGDEN AVENUE, SUITE 106

NAPERVILLE, ILLINOIS 60540

PHONE: (312) 369-6488

CLOSEOUT SURVEY REPORT

Performed For: *Cole HOSPITAL*

Date Performed: *6-11-85*

Instruments Used: *Survey Meter + Seale Well Counter Model 8725*

Data: * See attached diagram for area location in reference

Survey Location

Rm. No.	*Area
---------	-------

Results - mR/hr

Background = *.03 mR/hr*

1 = *0 mR/hr*
2 = *.05 mR/hr*
3 = *0 "*
4 = *.01 "*
5 = *0 "*
6 = *0 "*
7 = *0 "*
8 = *.02 "*
9 = *0 "*
10 = *0 "*
11 = *0 "*

Wipes

Results - CPM

Background = *25 cpm*

1 = *0 cpm*
2 = *3 cpm*
3 = *0 "*
4 = *0 "*
5 = *0 "*
6 = *1 "*
7 = *0 "*
8 = *2 cpm*
9 = *0 "*
10 = *0 "*
11 = *0 "*

Results:

☒ Results show no evidence of contamination

☐ Results show evidence of contamination; see comments

☐ See comments


Comments:

Performed by:

Ronald D. Edwards

Ronald D. Edwards,
Health Physicist

Room Utilized For Nuclear Medicine.
Outside Wall



readings m/hr