

APPENDIX A

NOTICE OF VIOLATION

Metropolitan Radiology Associates, Chartered
2139 Georgia Ave., N.W. Suite 4B
Washington, D.C. 20060

Docket No. 030-20741
License No. 08-20702-01

As a result of the inspection conducted on March 28, 1985 and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

Condition 16 of License No. 08-20702-01 requires that licensed material be possessed and used in accordance with the statements, representations and procedures contained in an application dated April 12, 1984, and a letter dated June 29, 1984.

1. Item 10 of the application dated April 12, 1984 requires that dose calibrator testing be performed in accordance with the procedures outlined on Page 9 of the attachment to the application.

Page 9 of the attachment requires that when daily constancy tests and quarterly linearity tests differ by greater than plus or minus five percent ($\pm 5\%$), instruments be repaired.

Contrary to the above, as of March 28, 1985, the readings were recorded, but not compared to the expected reading to identify a difference of greater than plus or minus five percent ($\pm 5\%$).

2. Page 15 of the application dated April 12, 1984 and Part 5 of the letter dated June 29, 1984 requires that wipe tests be performed on designated laboratory areas, and the results evaluated for contamination. Corrective action will be taken for any surface showing contamination greater than 200 disintegrations per minute per 100 square centimeters.

Contrary to the above, as of March 29, 1985, wipe tests were not adequately evaluated for contamination in that the counting equipment used for the evaluation did not have the sensitivity required to detect 200 disintegrations per 100 square centimeters.

This is a Severity Level IV violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Metropolitan Radiology Associates, Chartered, is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

8506030542 850528
REG1 LIC30
08-20702-01 PDR

OFFICIAL RECORD COPY

DL METRO RADIOLOGY - 0003.0.0
05/15/85