

JUN 28 1985

License No. 37-16144-01
Docket No. 030-10490
Control No. 103689

Messiah College
ATTN: Benjamin W. Myers, Ph.D.
Associate Professor of Biology
Grantham, Pennsylvania 17027

Gentlemen:

Please find enclosed the renewal of your NRC Material License.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the Region I Material Licensing Section, (215) 337-5239, so that we can provide appropriate corrections and answers.

Please be advised that you must conduct your program involving licensed radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, please note the items in the enclosed, "Requirements for Materials Licensees."

Since serious consequences to employees and the public can result from failure to comply with NRC requirements, the NRC expects licensees to pay meticulous attention to detail and to achieve the high standard of compliance which the NRC expects of its licensees.

You will be periodically inspected by NRC. A fee may be charged for inspections in accordance with 10 CFR Part 170. Failure to conduct your program safely and in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC will result in prompt and vigorous enforcement action against you. This could include issuance of a notice of violation, or in case of serious violations, an imposition of a civil penalty or an order suspending, modifying or revoking your license as specified in the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

OFFICIAL RECORD COPY

37-16144-01/LTR - 0001.0.0
06/24/85

8507260630 850628
REG1 LIC30
37-16144-01 PDR

ML10

We wish you success in operating a safe and effective licensed program.

Sincerely,

Original Signed By:
John E. Glenn

John E. Glenn, Chief
Nuclear Materials Safety Section B
Division of Radiation Safety
and Safeguards

Enclosures:

1. Amendment No. 03
2. Requirements for Materials Licensees

RI:DRSS
Joustra/mlb
6/28/85

RI:DRSS
Glenn
6/28/85

OFFICIAL RECORD COPY

37-16144-01/LTR - 0002.0.0
06/24/85

June 4, 1985

Region I
U.S. Nuclear Regulatory Commission
Attention: John E. Glenn, Ph.D., Chief
Nuclear Materials Section B
631 Park Avenue
King of Prussia, PA 19406

Dear Dr. Glenn:

This is in response to your letter dated May 24, 1985 concerning notice of violation for license number 37-16144-01, docket number 030-10490.

The violation came about due to an error in recording which of two gas chromatographs contained the radioactive source. The two gas chromatographs sit beside each other and someone took the wrong model information when the original license application was submitted. The Hewlett-Packard Model 18713A does not contain an electron capture cell, but the Micro Tek model MT 220 does.

I would therefore like to amend our license to delete the Hewlett-Packard Model 18713A and substitute Micro Tek model MT 220 with a Tracor electron capture detector model #1110192001, Serial #4501. This only alters item seven of amendment no. 2 of Material License dated December 6, 1984.

If the above amendment is approved, I would also like to amend item 15 of conditions to reflect the longer interval as suggested by the Tracor Instrument Notice which I sent to you with letter dated April 11, 1985.

Sincerely,

Benjamin W. Myers
Benjamin W. Myers, Ph.D.
Associate Professor of Biology

BWM/ms

cc: Dr. Noel Falk, Department Chairman
Dr. Karl Oberholser

7G/52

~~8507170662~~
2pp.

03689

"OFFICIAL RECORD COPY"

ML10

5/7/85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME

11:00

☒ A.M.
☐ P.M.

☐ INCOMING CALL

☒ OUTGOING CALL

☐ VISIT

PERSON CALLING

J.A. Soutra

OFFICE/ADDRESS

Region I

PHONE NUMBER

EXTENSION

5257

PERSON CALLED

Benjamin Myers

OFFICE/ADDRESS

Messiah College

PHONE NUMBER

EXTENSION

70
766-2511

CONVERSATION

SUBJECT

SUMMARY

Mr. Myers request for amendment addressed a change in leak test interval from 6 months to 3 yr. Mr. Myers refers to Tracor G.C., Messiah College license states only possession is a Hewlett Packard G.C.

We will send notice of violation with paragraph addressing ^{another} a request for amendment to add micro tech MT-220 Tracor G.C. then we will amend license to change leak test requirement.

"OFFICIAL RECORD COPY"

ML10

REFERRED TO:

ACTION REQUESTED

Cannot amend until NOV answered and new amend. request to add GC.

☐ ADVISE ME OF ACTION TAKEN.

INITIALS

DATE

ACTION TAKEN

INITIALS

DATE