

File

MAY 10 1985

Bates County Memorial Hospital
Nuclear Medicine Department
ATTN: James W. Baldwin
Administrator
615 W. Nursery Street
Butler, MO 64730

Gentlemen:

Enclosed is Amendment No. 02 renewing your NRC License No. 24-18740-01 in accordance with your request.

Please note that 10 CFR Part 35 has been revised effective October 12, 1982. This revision requires a representative of the nursing staff to be on the Medical Isotopes Committee. In view of this new requirement, you should appoint a member of your nursing staff to your Medical Isotopes Committee immediately. Also, note that we have added Condition No. 15. which requires a monthly xenon trap check in accordance with procedures in application dated February 22, 1985.

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.

8506040335 850510
REG3 LIC30
24-18740-01 PDR

MAY 10 1985

6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By
John R. Madera
Materials Licensing Section

Enclosure(s): Amendment No. 02

RIII

Madera/cm
05/08/85

CONVERSATION RECORD

TIME

202 (P)

DATE

5/7/85

TYPE

☐ VISIT

☐ CONFERENCE

☒ TELEPHONE

☐ INCOMING

☒ OUTGOING

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

ORGANIZATION (Office, dept., bureau, etc.)

TELEPHONE NO:

Jerry L. Harris

Nuclear Diagnosis

417/831-6520

Incs for Bates County Mem. Hosp.

SUBJECT

Delivery systems for Xe-133 @ Nevada City Hospital. What type?

SUMMARY

According to Jerry L. Harris, consultant, (060-133) Xe-133 rebreathing system from Atomic products is being used at Nevada City Hospital.

ACTION REQUIRED

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

JR-Maden

5/7/85

ACTION TAKEN

SIGNATURE

TITLE

DATE

50271-101

GPO : 1981 O - 361-526 (7227)

CONVERSATION RECORD

OPTIONAL FORM 271 (12-76)
DEPARTMENT OF DEFENSE