

GREENBERG RADIOLOGY CLINIC

1160 PARK AVENUE WEST, SUITE 2E

HIGHLAND PARK, ILLINOIS 60035

TELEPHONE (312) 433-0500

IRVING M. GREENBERG, M.D.
Diplomate American Board of Radiology
Diplomate American Board of Nuclear Medicine

BRENT M. GREENBERG, M.D.
Diplomate American Board of Radiology

MARK GREENBERG, M.D.
Diplomate American Board of Radiology

February 26, 1985

Bruce Mallett, PhD., Chief
Licensing Section U. S. Nuclear Regulatory Commission
799 Roosevelt Road
Glen Ellen, Illinois 60137

Re: #12-01211-07

Dear Dr. Mallett,

We request that you renew our Nuclear Regulatory Commission byproduct materials license. Enclosed, please find a check for \$580 to cover the renewal fee.

Our Nuclear Medicine program, as well as the authorized users listed on the current licence will remain the same.

Should you have any questions or require any information regarding this renewal request, please contact Mr. Ronald D. Edwards, Physicist of Radiation Protection Consultants, Ltd., Aurora, Illinois, who is authorized to answer any questions or provide any information you may request regarding this renewal or any amendment requests for the future.

Thank you in advance for your cooperation in this matter.

Very truly yours,

Mark Greenberg
Mark Greenberg, M.D.

Enclosure

mg:lk

Applicant	MAW 6 III
Check No.	6000 \$580
Amount Fee Category	75
Type	Ren
Date Check Rec'd	3/8/85
Received By	3/8/85

RECEIVED
FEB 27 1985
REGION III

8507150304 850621
REG3 LIC30
12-01211-07 PDR

CONTROL NO. 7 8 3 9 5

IRVING M. GREENBERG, M.D.
JOEL F. SAMMET, M.D.
BRENT M. GREENBERG, M.D.

TELEPHONE 433-0500

DRS. GREENBERG, SAMMET AND GREENBERG

RADIOLOGY

1160 PARK AVENUE WEST • SUITE 2E
HIGHLAND PARK, ILLINOIS 60035

February 13, 1981

John Cooper PhD.
Nuclear Regulatory Commission,
799 Roosevelt Rd.
Glen Ellyn, Ill 60137

Ref : 12-01211-07

Dear Mr. Cooper :

03 0-14085

Please amend our NRC ByProduct Materials License
to include the following changes ;

1. Delete the radiation survey instrument
in the original application - Victoreen 498.
2. Add the following survey instruments low
level Nuclear Associates 0-50 MR range
High level Victoreen CDV 715 0-5R range.
3. Our dose calibrator in a Radex Mark V.
4. Daily function checks will be performed
only on the days that the unit will be
used. Linearity checks will be performed on a semi-
annual basis. Annual calibration will be performed
using NBS traceable standards of Cobalt 57,
Cesium 137 and Barium 133.

The \$40.00 amendment fee is enclosed. Should you
have any questions please contact Mr Ronald Edwards
of Radiation Protection Consultants, Chicago, Ill 282-1888.

Thank you for your cooperation.

Very truly yours,
Brent M. Greenberg M.D.

BMG/be
2-13-81

~~81843859~~
1P

FEB 18 1981

Control No. 04438