



Industrial & Environmental Analysts, Inc.
P.O. Box 626 • Essex Junction, Vermont 05452 • 802-878-5138

4 June 1985

MS 16
P7

John D. Kinneman
Chief Nuclear Materials Section A
Division of Engineering and Technical Programs
NRC Region
631 Park Ave.
King of Prussia, Pennsylvania 19406

Control #10372

Dear Mr. Kinneman:

As per a telephone conversation with Mr. Tom Thompson concerning NRC license for an electron capture detector (Control #10372), we are transmitting herein the information which has been requested.

On May 31, 1985, IEA ordered a Shimadzu GC9APFE equipped with a shimdazu Nickle 63-10 millicurie electron capture detector. This instrument shall be installed in the IEA office-laboratory building which is not only locked but has a silent security system which is always on when the building is unattended.

The person responsible for taking the 6-month leak test will be Charlotte D. Sherman. The test will be run by Nuclear Sources & Services, Inc., P.O. Box 34042, Houston, Texas 77034. In the event that IEA needs to dispose of the electron capture, Shimadzu Scientific Instruments, Inc. will take the detector back.

If you need any further information from IEA prior to our obtaining a current NRC license please contact me at 802-878-5138. Thank you for your time and help in processing our request.

Sincerely,

Cathy Cutting

Cathy Cutting
Staff Scientist

8507160024 850626
REG1 LIC30
44-19169-02 PDR

21 JUN 85 9-12 PM '85

03726

RECEIVED-REGION

"OFFICIAL RECORD COPY"

Offices and laboratories located in: Essex Junction, Vermont
Research Triangle Park, North Carolina

ML10

JUN 06 1985

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME

☐ A.M.
☐ P.M.☐ INCOMING CALL☒ OUTGOING CALL☐ VISIT

PERSON CALLING

Thompson

OFFICE/ADDRESS

RT

PHONE NUMBER

EXTENSION

5303

PERSON CALLED

C. Cutting

OFFICE/ADDRESS

Industrial & Env. Analysts
Inc.

PHONE NUMBER

EXTENSION

(802) 878-5138

CONVERSATION

SUBJECT

Deficiencies in application for New Lic.

SUMMARY

Please describe C. Sherman's training to use G.C.

1) Please confirm you will secure G.C. when not attended.

2) If G.C. is some author on 44-19169-01 then you must reference these model numbers.

3) Reference or provide leak test info. Frequency and who will perform and analyze

4) Reference or state what you will do if you must dispose source

REFERRED TO:

ACTION REQUESTED

"OFFICIAL RECORD COPY"

ML10

☐ ADVISE ME OF ACTION TAKEN.

INITIALS

DATE

ACTION TAKEN

INITIALS

DATE