

Earl R. Gilreath
President

Application *APD 8-11*
Check No. *7842 \$40+*
Amount *1.2.56.36 \$80*
Type of *7 2 and*
Date Check Made *4/30/85*
Received by *CP*



March 20, 1985

Bruce Mallet, Ph.D., Chief
United States Nuclear Regulatory Commission
Region III
Material Licensing Section
799 Roosevelt Road
Glen Allen, Illinois 60137

RECEIVED
85 APR -1 AM 1:30

U.S. N.R.C.
LIC. FEE MGMT. BRANCH

RE: Amendment to License #34-13995-01

Dear Dr. Mallet:

Please amend license #34-13995-01 to:

- 1) Specifically authorize the possession of Sr-90 PTW source model #30-657, S/N 751327 from Nuclear Associates. The source is not to exceed 0.9mCi and is to be used for the sole purpose as a calibration device for consistency checks.
- 2) Increase the limit of possession of the byproduct material listed in Group VI of Schedule A, Section 35.100 of 10CFR35, from a total of one curie to a total of two curies of any sealed sources listed in Group VI of Schedule A, Section 35.100 of 10CFR35.
- 3) Delete the name of John Looper, M.D., from the license as an authorized user of byproduct material listed in Group VI of Schedule A, Section 35.100 of 10CFR35.

Enclosed is a check for \$40.00 as the amendment filing fee.

Sincerely yours,

Earl R. Gilreath

encl.

RECEIVED

MAR 25 1985

CONTROL NO. 78587
REGION III

cc: Vincent Seiwert, M.D., Chairman, Radiation Safety Committee
N.B.J. Tannous, Ph.D., Radiation Safety Officer

MAR 25 1985

8506050506 850510
REG3 LIC30
34-13995-01 PDR