



# farmington medical arts x-ray

23800 orchard lake road • farmington hills, michigan 48024 • telephone: 471-0675

May 17, 1985

LIC # 21-24380-01

030-18558

United States Nuclear  
Regulatory Commission  
Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

Dear Mr. Axelson:

This is in reponse to your letter dated May 3, 1985 referring to the routine safety inspection conducted by S. R. Lasuk at Farmington Medical Arts facility.

The following actions have been taken to correct the items of non-compliance:


- 1) Linearity tests of dose calibrator will be conducted on a quarterly basis. A linearity test was performed from 4/15/85 to 4/18/85 and found to be within +5% linearity!
- 2) Our local pharmacy, Syncor, Inc., will provide us with a leak test service of sealed reference sources on a quarterly basis. We will receive a written statement of each leak test, performed by their radiation safety office. The sources were leak tested approximately two weeks after the inspection date and no removable contamination was found on the test samples.

Quarterly inspections of all records, at all authorized places of use, by the radiation safety officer will prevent future noncompliances.

The Farmington Medical Arts facility is in full compliance as of this date.

The corrective actions stated above have been discussed with my staff and activities at all authorized places of use are conducted in compliance with requirements of the licensed program.

Sincerely,

  
Michael Lala, M.D.

ML:ean

8506050550 850517  
REG3 LIC30  
21-24380-01 PDR

IE07  
40  
MAY 20 1985