



DAVID E. HICKEY
DIRECTOR

BOARD OF LIGHT AND POWER

CITY OF MARQUETTE
2200 WRIGHT STREET
MARQUETTE, MI 49855-1398

PHONE 906-228-0320

September 4, 1996

U.S. Nuclear Regulatory Commission
Document Control Desk
Washington, DC 20555

To Whom It May Concern:

Pursuant to the Notice of Violation, License No. 21-20174-01; Document No. 030-18052, which we received on August 20, 1996 please find enclosed our Reply to a Notice of Violation.

If you have any further comments or questions regarding this Reply, please direct them to my attention.

Very truly yours,

William C. Pyle, P.E.,
Radiation Safety Officer

Enclosure: Reply to a Notice of Violation

c: David E. Hickey,
Executive Director

cc: Regional Administrator ✓
Region III
801 Warrenville Road
Lisle, IL 60532-4351

9610040021 960930
PDR ADDCK 03018052
C PDR

★ 1889 ★
100 YEARS OF PUBLIC SERVICE
★ 1989 ★

SEP 06 1996

September 4, 1996

U.S. Nuclear Regulatory Commission
Document Control Desk
Washington, D.C. 20555

Re: REPLY TO A NOTICE OF VIOLATION

(1) The gauge inventory program was set up with the assistance of the architect/engineer and gauge manufacturer at the time of installation. If the requirement to operate the shutters during the inventory was discussed at that time, it was overlooked and not incorporated into the final inventory procedure. (2) The inventory procedure has been modified so that the shutters are operated at each inventory and the condition logged. (3) No further steps are planned to be taken. (4) The revised inventory procedure was implemented on July 18.

cc: Regional Administrator, Region III