

OCT 2, 1996

Mr. John D. Kinneman  
US NUCLEAR REGULATORY COMMISSION  
REGION 1  
475 ALLENDALE ROAD  
KING OF PRUSSIA, PA 19406-1415

**REPLY TO A NOTICE OF VIOLATION**

Dear Mr. Kinneman:

Please find enclosed my response to the recent violations documented at the August 5th inspection.

The following violations will be corrected by Nov 1, 1996.

A.1. The original NRC application was set up to accommodate either Vithel Jathar or Craig Sockol as the radiation safety officer of repromedix. When Vithel Jathar left, Craig Sockol took over the responsibility as RSO. It was assumed that notification was not required, as Craig Sockol has had past experience as RSO at Immunotech Corp and is now the official RSO of repromedix.

A. 2. The technician was trained on-site verbally and the documentation was lost. The technician will be scheduled for a complete radiation safety training session for handling licensed materials.

B. An annual audit will be scheduled and documented for compliance. The company did not believe an audit was required for its minimal usage.

C. Some radiation signs were mistakenly taken down. Radiation signs will be posted in the lab area and in the adjacent storage room.


D. It was assumed that the quantity received was too low to warrant an outer box wipe test. All future non-exempted radioactive packages will be wipe tested and documented.

**RETURN ORIGINAL TO  
REGION I**

9610290032 961021  
PDR ADOCK 03033462  
C PDR

E. It was assumed that the minimal usage of radioactivity was well below the allowable limits to into the sewer system. A complete accounting of liquid radioactive waste disposal in the sewer system is now being recorded for unity.

Yours Truly

A handwritten signature in cursive script, reading "Craig S. Sockol".

Craig Sockol  
President