

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s); and to import such byproduct and source material. This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	
1. Prinkki and Associates	3. License number 25-23175-01
2. 3425 Lloyd Court Missoula, Montana 59803	4. Expiration date May 31, 1990
	5. Docket or Reference No. 030-20341
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form
A. Cesium-137	A. Sealed sources (Troxler Dwg. No. A-102112)
B. Americium-241	B. Sealed neutron sources (Troxler Dwg. No. A-102451)
C. Cesium-137	C. Sealed sources (Campbell Pacific Nuclear Model CPN-131)
D. Americium-241	D. Sealed neutron sources (Campbell Pacific Nuclear Model CPN-131)
	8. Maximum amount that licensee may possess at any one time under this license
	A. Not to exceed 10 millicuries per source
	B. Not to exceed 50 millicuries per source
	C. Not to exceed 10 millicuries per source
	D. Not to exceed 50 millicuries per source

8507170313 850502
REG4 LIC30
25-23175-01 PDR

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
25-23175-01
Docket or Reference number
030-20341

E. Cesium-137/
Americium-241

E. Combined sealed sources
(Troxler Dwg. No.
A-100281, Rev. B)

E. Not to exceed
10 millicuries
of Cesium-137
and 50 milli-
curies of
Americium-241
per source

9. Authorized use

- A. and B. For use in Troxler Model 3400 Series moisture/density gauges for measurement of properties of materials.
- C. and D. For use in Campbell Pacific Nuclear Model B(R) and Model MC Series moisture/density gauges for measurement of properties of materials.
- E. For use in Troxler Model 2401, 2402, 2451, or 2452 moisture/density gauges for measurement of properties of materials.

CONDITIONS

- 10. Licensed material may be used at 3425 Lloyd Court, Missoula, Montana, and at temporary job sites of the licensee anywhere in the United States where the U. S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
- 11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
25-23175-01

Docket or Reference number
030-20341

12. Licensed material shall be used by, or under the supervision and in the physical presence of licensee employees who have completed either the Troxler or Campbell Pacific Nuclear training course in the use of the devices and have been designated by Larry Prinkki.
13. A. (1) Each sealed source containing licensed material shall be tested for leakage and/or contamination at intervals not to exceed 6 months. In the absence of a certificate from a transferor indicating that a test has been made within 6 months prior to the transfer, a sealed source received from another person shall not be put into use until tested.
- (2) The periodic leak test required by this condition does not apply to sealed sources that are stored and not being used. The sources excepted from this test shall be tested for leakage prior to any use or transfer to another person unless they have been leak tested within 6 months prior to the date of use or transfer.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the test with the U. S. Nuclear Regulatory Commission, Region IV, 611 Ryan Plaza Dr., Suite 1000, Arlington, Texas 76011, describing the equipment involved, the test results, and the corrective action taken.

MATERIALS LICENSE
SUPPLEMENTARY SHEETLicense number
25-23175-01
Docket or Reference number
030-20341

13. (continued)

- D. The licensee is authorized to collect leak test samples in accordance with the procedures described in the licensee's application dated April 8, 1985, for analysis by Troxler Electronic Laboratories. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.
14. Sealed sources containing licensed material shall not be opened or removed from moisture/density gauges by the licensee.
15. Maintenance or repair of portable devices involving removal of the sealed sources from the devices or removal or dismantling of shielding may be performed only by the device manufacturer or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
16. The licensee shall conduct a physical inventory every 6 months to account for all sealed sources received and possessed under the license. The records of the inventories shall be maintained for 2 years from the date of the inventory for inspection by the Commission, and shall include the quantities and kinds of byproduct material, manufacturers' names and model numbers, location of sealed sources and the date of the inventory.
17. The licensee may transport licensed material or deliver licensed material to a carrier for transport in accordance with the provisions of Title 10, Code of Federal Regulations, Part 71, "Packaging and Transportation of Radioactive Material."
18. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated April 8, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date MAY 2 1986

By *Original Signed By*
Q. J. Cain
Nuclear Materials Safety Section
Region IV
Arlington, Texas 76011