



Veterans
Administration

June 13, 1985



U.S. Nuclear Regulatory Commission, Region I
631 Park Avenue
King of Prussia, PA 19406

ATTN: John E. Glenn, Chief
Nuclear Materials Safety Section B
Division of Radiation Safety and Safeguards

License No. 20-11750-02
Docket No. 030-08433

Subject: Inspection Report No. 85-01

This is in response to your letter dated June 3, 1985
regarding the information on AC/DC rate meter.

We have placed an order for the following two rate meters
from the Ludlum Company on June 5, 1985. We expect to
receive them by July 5, 1985.

1. Alarm rate meter (Ludlum) with beta-gamma end window
2. Alarm rate meter (Ludlum) with pancake probe

Additionally, on May 13, 1985 we requested an amendment to
an item in Appendix G, "General Rules for Safe Use of
Radioactive Material" of our license, from "monitor hands
and clothing for contamination after each procedure or
before leaving the area", to "monitor hands and clothing
for contamination after each procedure or before leaving
the radioisotope area".

THOMAS J. STRANOVA
Medical Center Director

8507170368 850712
REQ1 LIC30
20-11750-02 PDR