

MAY 24 1985

Docket No. 30-06097

License No. 37-03330-04

United States Steel Corporation  
ATTN: G. P. Huffman, Chairman  
USS Technical Center  
Radiation Committee  
One Tech Center Drive  
Monroeville, Pennsylvania 15146

Gentlemen:

Subject: Inspection Report No. 030-06097/85-01

This refers to your letter dated March 22, 1985, in response to our letter dated February 22, 1985.

Thank you for informing us of the corrective and preventive actions documented in your letter. These actions will be examined during a future inspection of your licensed program.

The following is provided for your information. Smears for removable contamination were taken during the inspection with the following results: (1) can in which old accelerator target is stored: 21,400 dpm tritium/wipe or approximately 10,000 dpm/100 cm<sup>2</sup>; (2) outside of target area of old accelerator: 9,700 dpm tritium/wipe or approximately 4,800 dpm/100 cm<sup>2</sup>; and, (3) outside of target area of new accelerator, negative results for removable activity. This information should be used to evaluate the adequacy of current postings and safety procedures in this room.

Your cooperation with us is appreciated.

Sincerely,

Original Signed By:  
John D. Kinneman

John D. Kinneman, Chief  
Nuclear Materials Section A  
Nuclear Materials Safety & Safeguards Branch  
Division of Radiation Safety & Safeguards

cc:  
Public Document Room (PDR)  
Nuclear Safety Information Center (NSIC)  
Commonwealth of Pennsylvania

8506070145 850524  
REQ1 LIC30 PDR  
37-03330-04

OFFICIAL RECORD COPY

RL U STEEL 85-01 - 0001.0.0  
05/17/85

RETURN ORIGINAL TO  
REGION I

IE:07

MAY 24 1985


United States Steel Corporation

2

bcc:  
Region I Docket Room (w/concurrences)

  
RI:DRSS  
McFadden/djh

05/21/85

  
RI:DRSS  
Kinneman

05/21/85  
OFFICIAL RECORD COPY

RL U STEEL 85-01 - 0002.0.0  
05/17/85