



Mr. A. Bill Beach
Regional Administrator, Region III
United States Nuclear Regulatory
Commission
801 Warrenville Road
Lisle, IL 60532-4351

October 7, 1996

Re: Confirmatory Action Letter (CAL)
CAL No. RIII-96-011
Docket No. 030-33731
License No. 34-18214-03E

Dear Mr. Beach:

In reference to the telephone conversation between Mr. D. Wiedeman and the undersigned on Friday, October 4, 1996, and in reference to your letter of August 19, 1996, as well as our response to this letter, as well as reference to all referenced letters and correspondence, please be advised that we offer further clarifications as given below.

In reference to the paragraph in your letter of August 19, 1996,

1. I have indeed reviewed the license no. 34-18214-03E and the referenced documents, conditions, pertinent rules and have gone over this material with the following persons:
 - i. Ms. Isabel Overhoff, Vice President and Head of Administration
 - ii. Mr. Dell Williamson, Q/A Manager
 - iii. Mr. Mark Atanas, Manager of Mechanical Assembly Department

In a meeting with this group of people we have discussed every item to ensure that everyone has been advised of the requirements.

2. I also wish to advise you that the documents to be posted have in fact been posted, both in the assembly area, as well as near our shipping/receiving location. As mentioned above, Mr. Williamson and Ms. Overhoff, who helped draft these documents themselves, have again reviewed this material.
3. Management Audits, as required, will be performed as indicated. Mr. Ken Fritz, consulting RSO to our Company, has been advised, has been given copies of the pertinent correspondence, and has agreed to perform this duty.
4. I reconfirm that only two ^3H calibrators were transferred to our customers, S/N 1919, S/N 1939. I have reviewed calculations and have concluded, once more, that, due to oversight, these lecture bottles did contain about 9 mCi of ^3H , which exceeds the allowable limit of 1 mCi.

9610300088 961025
PDR ADDOCK 03033371
C PDR

P.O. BOX 182, 1160 U.S. ROUTE 50, MILFORD, OHIO, 45150-9705, U.S.A.
TELEPHONE (513) 248-2400, FACSIMILE (513) 248-2402, E-MAIL OVERHOFF@IX.NETCOM.COM

OCT 08 1996

US NRC, Region III
Page 2
October 7, 1996


Re: Confirmation Action Letter (CAL)
CAL No. RIII-96-011

5. See above. These referenced written procedures are the documents which have been posted so that every one concerned will be reminded of the requirements as time goes by.
6. Please note that the above cited documents clearly state that the lecture bottle shall be detached from the rest of the calibrator prior to packing and shipping. This fact was again stressed during the meeting with our staff members as mentioned in paragraph 1.

I hope that this confirmation will help to clear up everything. If you need any further information, please do not hesitate to let me know.

Yours Sincerely,

MWO:blw


M. W. Overhoff, PhD
President and RSO