

# SAFETY INSPECTION

Page 1 of \_\_\_\_\_

1. LICENSEE

Med. Share Inc.  
31700 W. 13 mile Rd  
Suite 200  
PARDINGTON HILLS, MI 48334

2. REGIONAL OFFICE

REGION III  
U S NUCLEAR REGULATORY COMMISSION  
801 WARRENVILLE ROAD  
LISLE IL 60532-4351

3. DOCKET NUMBER(S)

030-31441

4. LICENSE NUMBER(S)

21-26098-01

5. DATE OF INSPECTION

9/19/96

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- ☒ 1. Within the scope of this inspection, no violations were observed.
- ☐ 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.
- ☐ 3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a **NOTICE OF VIOLATION**, which is required to be posted in accordance with 10 CFR 19.11.
- ☐ A. \_\_\_\_\_ was not properly posted to indicate the presence of a \_\_\_\_\_ 10 CFR 20.203(b),(c),(d),(e) or 34.42.
- ☐ B. \_\_\_\_\_ of sealed sources were not performed at the proper frequencies. 10 CFR \_\_\_\_\_ or License Condition Number \_\_\_\_\_
- ☐ C. Records of \_\_\_\_\_ were not properly maintained. 10 CFR \_\_\_\_\_ or License Condition Number \_\_\_\_\_
- ☐ D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.
- ☐ E. Reports or notification of \_\_\_\_\_ were not made in accordance with 10 CFR \_\_\_\_\_ or License Condition Number \_\_\_\_\_
- ☐ F. \_\_\_\_\_

300026

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE -- LICENSEE

9610300128 960919  
PDR ADOCK 03031441  
C PDR

DATE

SIGNATURE -- NRC INSPECTOR

*[Signature]*

DATE

9/19/96

# SAFETY INSPECTION

Page 2 of \_\_\_\_\_

1. LICENSEE  
*Med Share INC.  
31700W 13 mile Rd  
Suite 200  
Farmington Hills MI 48334*

2. REGIONAL OFFICE  
REGION III  
U S NUCLEAR REGULATORY COMMISSION  
801 WARRENVILLE ROAD  
LISLE IL 60532-4351

3. DOCKET NUMBER(S)  
*030-31441*

4. LICENSE NUMBER(S)  
*21-26098-01*

5. DATE OF INSPECTION  
*9/19/96*

3. (Continued)

<input type="checkbox"/>	G.	
<input type="checkbox"/>	H.	
<input type="checkbox"/>	I.	

☒ 4. The violations listed below are not being cited because they were self-identified, and corrective action was or is being taken, and the remaining criteria in 10 CFR 2, App. C, to exercise discretion were satisfied.

- ☒ A. *10CFR35.70 requires that surveys for removable contamination be performed weekly whenever radioactive material is used. Contrary to this, no such tests were performed at Jackson & Treador for weeks -*
- ☐ B. *9/26, 10/3, 4/29 of 1995 (FNUIT)  
9/5, 9/28, 10/6 10/27, 4/6, 11/13 11/15 & 11/29*
- ☐ C. *30399222*
- Corrective Actions - Self Identified by RSO & MPC.  
- The Technologist was fired  
- Lesson learned was covered during quarterly meeting on 12/6/95*