

SAFETY INSPECTION

Page 1 of 2

1. LICENSEE

CENTRAL MICHIGAN UNIVERSITY

2. REGIONAL OFFICE

REGION III
U S NUCLEAR REGULATORY COMMISSION
801 WARRENVILLE ROAD
LISLE IL 60532-4351

3. DOCKET NUMBER(S)

license

21-01432-03

4. LICENSE NUMBER(S)

bokeh

030-00811

5. DATE OF INSPECTION

10-17-96

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- ☐ 1. Within the scope of this inspection, no violations were observed.
- ☒ 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on those actions at this time.
- ☒ 3. During this inspection certain of your activities, as described below or attached, were in violation of NRC requirements. This form is a **NOTICE OF VIOLATION**, which is required to be posted in accordance with 10 CFR 19.11.

☐ A. _____ was not properly posted to indicate the presence of a _____ 10 CFR 20.203(b),(c),(d),(e) or 34.42.

☐ B. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ or License Condition Number _____

☒ C. Records of *Leak Tests AND INSTRUMENT CALIBRATIONS* were not properly maintained. 10 CFR *19.11* or License Condition Number *19A*

☐ D. Documents were not properly posted or otherwise made available. 10 CFR 19.11.

☐ E. Reports or notification of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____

☒ F. *LICENSE CONDITION 19A; ITEM 10 OF THE APPLICATION DATED 6/25/91, REQUIRES THE RADIATION SAFETY COMMITTEE TO MEET IN THE SPRING AND FALL OF EACH YEAR. CONTRARY TO THE ABOVE, THE COMMITTEE HAS NOT MET SINCE 1992.*

IE079

9610300152 961017
PDR ADOCK 03000811
C PDR

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

SIGNATURE - LICENSEE

DATE

SIGNATURE - NRC INSPECTOR

DATE

[Signature]

10/17/96

[Signature]

10-17-96

SAFETY INSPECTION

Page 2 of 2

1. LICENSEE

CENTRAL MICHIGAN UNIVERSITY

2. REGIONAL OFFICE

REGION III
U S NUCLEAR REGULATORY COMMISSION
801 WARRENVILLE ROAD
LISLE IL 60532-4351

3. DOCKET NUMBER(S)

030-00811

4. LICENSE NUMBER(S)

21-01432-02

5. DATE OF INSPECTION

10/17/96

3. (Continued)

☒ G.

L/C 19A and SECTION 10 of the Application dated 6/25/91
list the methods of instrument calibration used by
the University. Contrary to the above the licensee was
not calibrating the survey instruments using the
approved procedure, as of 10/17/96.

☐ I.

4. The violations listed below are not being cited because they were self-identified, and corrective action was or is being taken, and the remaining criteria in 10 CFR 2, App. C, to exercise discretion were satisfied.

☐ A.

☐ B.

☐ C.