

ENCLOSURE  
NOTICE OF VIOLATION

Wise Clinic

License No. 45-19154-01

The following violations were identified during an inspection conducted on March 19, 1985. The Severity Levels were assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

1. License Condition 15 requires that licensed material be used in accordance with statements and procedures contained in your application dated June 27, 1984. One procedure in your application requires that radiation level surveys be conducted daily.

Contrary to the above, radiation level surveys had been conducted on a weekly basis.

This is a Severity Level IV violation (Supplement VI).

2. License Condition 15 requires that licensed material be used in accordance with statements and procedures contained in your application dated June 27, 1984. One procedure in your application states that the inner container of radioactive shipments will be checked for contamination.

Contrary to the above, the licensee had not checked for contamination on the inner container of radioactive materials received at the clinic.

This is a Severity Level IV violation (Supplement VI).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: APR 02 1985

B507180370 B50402  
REG2 LIC30  
45-19154-01 PDR