

MAY 24 1985

Docket No. : 030-18926  
License No.: 46-19584-01  
Control No.: 70182

V. A. Medical Center  
American Lake  
Tacoma, Washington 98493

Attention: Mr. William E. Claypool  
Medical Center Director

Gentlemen:

Enclosed is the NRC license amendment which you requested.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify us so that we can provide appropriate corrections and answers.

You should note that the fume hood used for iodinations should include a charcoal filter to ensure that the requirements of 10 CFR 20.106 are satisfied.

Any future correspondence relating to your license should specifically reference your license and docket numbers to expedite your inquiry.

Sincerely,

Beth A. Riedlinger  
Health Physicist (Licensing)  
Nuclear Materials Safety Section

Enclosure: Attachment A  
Amendment No. 6 to License No. 46-19584-01

8507180558 850524  
REG5 LIC30  
46-19584-01 PDR

OFFICE	RV... <i>BAK</i>					
SURNAME	Riedlinger:fr					
DATE	5/24/85					