

TABLE 1

U S NUCLEAR REGULATORY COMMISSION
 OFFICE OF INSPECTION AND ENFORCEMENT
 CONFIRMATORY MEASUREMENTS PROGRAM
 FACILITY: CALLAWAY
 FOR THE 2 QUARTER OF 1985

SAMPLE	ISOTOPE	-----NRC-----		----LICENSEE----		---LICENSEE:NRC---		
		RESULT	ERROR	RESULT	ERROR	RATIO	RES	T
L WASTE	H-3	2.9E-03	4.0E-05	3.1E-03	0.0E-01	1.1E 00	7.2E 01	A
	G BETA	4.0E-07	8.0E-08	4.8E-07	0.0E-01	1.2E 00	5.0E 00	A
	SR-89	1.5E-09	7.0E-09	1.5E-08	0.0E-01	1.0E 01	2.1E-01	N
	SR-90	0.0E-01	4.0E-09	2.3E-08	0.0E-01	0.0E-01	0.0E-01	N

T TEST RESULTS:

A=AGREEMENT

D=DISAGREEMENT

*=CRITERIA RELAXED

N=NO COMPARISON