

FILE COPY

MAY 22 1985

Martin Marietta Refractories Company
ATTN: C.E. Hoban
Plant Manager
755 Lime Road
Woodville, OH 43469

Gentlemen:

Enclosed is Amendment No. 03 to your NRC License No. 34-13543-01 in accordance with your request.

Please note that as currently written your license does not authorize you to install, relocate, maintain, repair or remove gauges. In order to obtain this authorization you must submit the following information:

1. Verify that Victoreen Instrument Company is licensed by the NRC to perform survey meter calibrations and that their calibrations are performed using a radioactive source.
2. Modify your survey meter calibration procedures to state that the meter will be calibrated by an NRC or Agreement State authorized company at least annually and following repair.
3. Identify which operations you wish to perform i.e., installation, relocation, maintenance, repair and/or removal.
4. Submit step by step radiation safety procedures to be followed in performing each of the above operations. These procedures should include, but not be limited to who, when and where surveys will be performed; evaluations of occupancy and radiation dose to employees, etc. Your procedures must be equivalent to those performed by the manufacturer, must assure the safety of the public and employees, and assure that pertinent regulations and license conditions are met.

Also, please note that the license does not authorize you to perform your own analysis of leak test samples. In order to obtain this authorization, you must demonstrate your method is sufficiently sensitive to detect 0.002 microcuries of cesium-137.

Information submitted in response to this letter should state that it is additional information and not a new amendment request.

8506100081 850522
REG3 LIC30
34-13543-01 PDR

MAY 22 1985

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.
6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

Martin Marietta Refractories Company 3

MAY 22 1985

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By
Evelyn R. Matson
Materials Licensing Section

Enclosure(s): Amendment No. 03

RIII

ERM 5/20
Matson/cm
05/17/85