

MASSAC MEMORIAL HOSPITAL

MEMORIAL HEIGHTS — METROPOLIS, ILLINOIS 62960

LOREN L. ERWIN
ADMINISTRATOR

3 June 1985

Materials Licensing Section
U.S. Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Gentlemen,

REFERENCE: License # 12-24313-01

Please accept this letter as our request to amend the above referenced Byproduct Materials License. The amendment requested for is the change in location of our Nuclear Medicine Department.

Please note that the location indicated in our license application has not been used for Nuclear Medicine. A check, issued to the U.S. Nuclear Regulatory Commission for One Hundred Twenty Dollars as amendment filing fee, is attached.

We hope that the information submitted is satisfactory.

Sincerely,

Loren L. Erwin

Loren L. Erwin
Administrator

Applicant	<i>June 9/11</i>
Check No.	<i>007504 \$120</i>
Amount/Fee Category	<i>EX 713</i>
Type of Fee	<i>Amel</i>
Check Rec'd	<i>6/13/85</i>
Received By	<i>[Signature]</i>

170.11(a)(a)
FEE EXEMPT #16

8507230529 850628
REG3 LIC30
12-24313-01 PDR

U.S. N.R.C.
METROPOLIS BRANCH

85 JUN 13 AM 1:49

RECEIVED
JUN 07 1985
REGION III

CONTROL NO. 7 913 2

JUN 7 1985

ASHWIN PATEL
Consultant In Radiation Safety & Quality Assurance
For Nuclear Medicine & Ultrasound

May 26, 1985

Mahendra I. Vyas, M.D.,
Radiology Director,
Massac Memorial Hospital,
Metropolis, IL 62960.

Dear Dr. Vyas,

As I indicated in my survey report, the NRC license should be amended to reflect the change in location of the nuclear medicine department. Please transcribe the following letter on the hospital stationery, signed by the administrator, and mail TWO copies of the letter, the enclosed sketch of the nuclear medicine department, and a check for One Hundred Twenty Dollars issued to the U.S. Nuclear Regulatory Commission, to the Licensing section of the NRC. Should you have any questions, please do not hesitate to contact me (I will be on vacation from today until June 11, and I will be out of State during that time period. Please note this amendment application should be submitted to the NRC without any delay:

DATE: _____

Materials Licensing Section,
U.S. Nuclear Regulatory Commission,
Region III,
799 Roosevelt Road,
Glen Ellyn, Illinois 60137.

Gentlemen,

REFERENCE: License NO. 12-24313-01

Please accept this letter as our request to amend the above referenced Byproduct Materials license. The amendment requested for is the change in location of our Nuclear Medicine Department.

Please note that the location indicated in our license application has not been used for Nuclear Medicine. A check, issued to the U.S. Nuclear Regulatory Commission for One Hundred Twenty Dollars as amendment filing fee, is attached.

We hope that the information submitted is satisfactory.

Sincerely,

Administrator

RECEIVED


JUN 07 1985

REGION III

CONTROL NO. 7 9 1 3 2

MASSAC MEMORIAL HOSP., METROPOLIS, IL.

LICENSE NO. 12-24313-01
IL-00581-01.



- A: ^{99m}Tc GENERATOR SHIELDED BY 2" THICK LEAD.
B: "LONGER" HALF-LIFE WASTE (^{131}I , ^{67}Ga , ETC.) SHIELDED BY $\frac{3}{4}$ " LEAD.
C: L-BLOCK: DOSE PREP AREA.
D: DOSE CALIBRATOR
E: ^{99m}Tc WASTE CONTAINER SHIELDED BY $\frac{3}{16}$ " LEAD.
F: SCINTILLATION CAMERA DETECTOR
G: CAMERA CONSOL & FORMATTER
H: SECURED DOOR.

ITEM # 11. (6/84)

CONTROL NO. 9132