

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

34-20317-01

Docket or Reference number

030-18243

Amendment No. 03

Henry R. Silverman, III, M.D.
3939 Monroe Street
Suite 123
Toledo, OH 43606

In accordance with letter dated May 25, 1985, License Number 34-20317-01 is amended as follows:

Condition 16. is amended to read:

16. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated September 8, 1982 (including ALARA Program); letters dated March 16, 1984, June 1, 1984 and May 25, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

Condition 17. is added:

17. The licensee is authorized to calibrate their radiation detection instrument(s) and dose calibrator at intervals not to exceed those outlined in Regulatory Guide 10.8, October 1980. The Radiation detection instrument(s) shall be calibrated by a company or person whose procedures have been approved and are on file with the NRC. The dose calibrator shall be calibrated by the licensee in accordance with Section 2 of Appendix D of Regulatory Guide 10.8, October 1980.

8507230554 850613
REG LIC30
34-20317-01 PDR

For the U.S. Nuclear Regulatory Commission

Date June 13, 1985

Original Signed
By George M. McCann
Materials Licensing Section, Region III

11 17230BK JUN 17 1985 COPY 2