

SOURCE CONTAINMENT AND DEVICES BRANCH

OFFICE OF NUCLEAR MATERIAL SAFETY AND SAFEGUARDS DIVISION OF INDUSTRIAL AND MEDICAL NUCLEAR SAFETY

3/28/96

FAX TRANSMITTAL

TO

FROM

1.

George Stoddard
FAX# () 201-263 0477

Seta Control

2.

FAX# ()

3.

FAX# ()

4.

FAX# ()

5.

FAX# ()

6.

FAX# ()

COMMENTS:

9611220236 961101
PDR RC *
SSD PDR

COVER SHEET PLUS 1 PAGES

FROM:

Michele Burgess

PHONE:

(301) 415-5868

FAX:

(301) 415-5369

9611220236 2PP

def call

March 28, 1996

Mr. Stoddard,

As you requested, here are the points that we discussed on the phone today. Please address the following:

1. Describe all design changes between the specifically licensed (SL) and the generally licensed (GL) devices.
2. Provide a description or drawing of the GL label and indicate its location on the device. Please verify that the GL label will not be present on the SL devices.
- ✓ 3. Verify that the label will have the date of assay for the GL devices.
4. List any services, maintenance or testing that you will provide.
5. List any servicing, maintenance or testing that you want the GL user to perform.
6. Please provide a copy of the users manual. We do not have a copy in the file, and we need to review the manual for GL users.
7. Provide changes and/or additions to the QA/QC procedures that will be performed for the GL device that will assure that all GL devices contain the appropriate design features and labeling, and that the user receives all required information.

I look forward to receiving your response.

Michele Burgess



betacontrol

P.O. Box 235 • 435 Route 202 • Towaco, N.J. 07082

Mr. Steve Baggett
U.S. Nuclear Regulatory
Commission
Mail Stop T8F5
Washington, DC 20555