

From: [Cender, Laura](#)
To: [Jackson, Alan \(RSO\)](#); [Doemer, Anthony](#)
Cc: [Bevins, Nicholas](#); [Macdonald, Erin](#); [Steiner, Joseph](#)
Subject: RE: HDR Corrections for: Pending Amendment for Henry Ford Hospital - Additional Information Requested
Date: Friday, April 10, 2020 7:39:00 AM

Thank you Alan, I have no further questions at this time.

-Laura

Laura Cender
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712

From: Jackson, Alan (RSO) <AlanJ@rad.hfh.edu>
Sent: Wednesday, April 08, 2020 3:28 PM
To: Cender, Laura <Laura.Cender@nrc.gov>; Doemer, Anthony <adoemer1@hfhs.org>
Cc: Bevins, Nicholas <nickb@rad.hfh.edu>; Macdonald, Erin <erinma@rad.hfh.edu>; Steiner, Joseph <josephst@rad.hfh.edu>
Subject: [External_Sender] HDR Corrections for: Pending Amendment for Henry Ford Hospital - Additional Information Requested

Laura,

I hope that the attached letter adequately addresses the concerns you had.

-Alan

Alan M. Jackson, MS. CHP
Radiation Safety Officer
Henry Ford Hospital
2799 West Grand Boulevard
Detroit, MI 48202
(734) 657-4133 Cell

From: [Cender, Laura](#)
To: [Jackson, Alan \(RSO\)](#)
Subject: Pending Amendment for Henry Ford Hospital - Additional Information Requested
Date: Tuesday, April 07, 2020 12:06:05 PM

Hello Alan,

I was completing my review of your amendment request and noted that the shielding report provided in Appendix A is inconsistent in describing the shielding is present.

Specifically, the wall material described in the "Room Description" section of the report differs from what is provided in the table on the following page. The differing descriptions are specific to the South and West wall, but I request that you provide a signed response confirming the actual wall material present at each reference location.

Thank you very much for your help.

Laura