

From: [Gryglak, Magdalena](#)
To: ["pbyrne@mpcphysics.com"](mailto:pbyrne@mpcphysics.com)
Subject: Request to add Dr. Lee as an AU and F-18 for calibration, NRC license no. 13-17956-01, Marion General Hospital
Date: Monday, March 30, 2020 8:58:00 AM

Good morning Mr. Byrne,

I have reviewed the licensee's request dated March 5, 2020.

Please provide the following information in support of the request:

1. Copy of Dr. Lee's medical license.
2. Describe the area/room at 1410 W Bella Dr., Marion, Indiana where Fluorine-18 (F-18) will be used for calibration.
3. Describe the equipment that will be calibrated with F-18.
4. Describe the process for calibrating the equipment.
5. Describe measures to reduce the unnecessary exposure to members of the public in the unrestricted areas (e.g. physical barriers in place, methods, duration, calculations etc.) to demonstrate that the dose limits in Title 10 of the Code of Federal Regulations Part 20 are not exceeded.
6. Provide copies of the Certified Nuclear Medicine Technologist (CNMT) cards for Jay Taylor and Rachel Delaney-Culp.
7. Describe Mr. Taylor's and Ms. Delaney-Culp's hands-on work experience (locations and duration of experience, types/form/quantity of material handled, activities performed).

Please provide your response in a signed (by Amy Droom) and dated letter by May 5, 2020. You may submit the response directly to me via email.

Please acknowledge the receipt of this email and let me know if you have any questions.

Thank you

Magdalena R. Gryglak
U.S. NRC Region III
630-829-9875