

DCD

December 10, 1996

Ms. Debra Rasper
Corporate Operating Officer
Community Hospitals of
Indiana, Inc.
1500 N. Ritter Avenue
Indianapolis, IN 46219

SUBJECT: NOTICE OF VIOLATION DATED OCTOBER 31, 1996

Dear Ms. Rasper:

This acknowledges receipt of your undated letter signed by Andrea Browne, Ph.D., in response to our letter dated October 31, 1996, transmitting a Notice of Violation.

We have reviewed your corrective actions, which appear to be adequate, and have no further questions at this time. These corrective actions will be examined during a future inspection.

Sincerely,

Original Signed by Roy J. Caniano

Cynthia D. Pederson, Director,
Division of Nuclear Materials Safety

License No.: 13-06009-01
Docket No.: 030-01625

cc: Andrea Browne, Ph.D.

bcc w/ltr undated: PUBLIC

DOCUMENT NAME: G:\LTRS2LIC\MTLS\030\96301625.L03

To receive a copy of this document, indicate in the box: "C" = Copy without enclosures "E" = Copy with enclosures "N" = No copy

OFFICE	DNMS/RIII	DNMS/RIII	DNMS/RIII						
NAME	DWIEDEMAN:jaw	JMADERA	CPEDERSON						
DATE	12/1/96	12/10/96	12/1/96						

OFFICIAL RECORD COPY

9612160037 961210
PDR ADOCK 03001625
C PDR

IE07



Community Hospitals Indianapolis

Community Hospital East

16th Street and Ritter Avenue
P.O. Box 19805
Indianapolis, Indiana 46219-0805
Telephone (317) 355-1411

Document Control Desk
USNRC
Washington, DC 20555

Reply to a Notice of Violation **Materials License 13-06009-01**

In response to **NRC Inspection Report 030-01625/96001 (DNMS)** and notice of violation received Nov. 4, 1996:

Reason for violation: An unusual dummy train was used for the first time by the planning physicist. When the planning physicist used the dummy a correction was not made to the specified length of the strand to account for the hub offset.

Corrective steps taken and results achieved: Record of all dummy strands owned will be placed in the QC manual and available for all individuals planning HDR treatments.

Corrective steps that will be taken to avoid future violations: Demonstration of ability to calculate the appropriate dummy length will be included in the competency check for all individuals performing HDR treatment planning.

Date when full compliance will be achieved: 8/17/96

Andrea D. Browne, Ph.D.
Radiation Safety Officer



#378465
01/07/2000

Printed on recycled paper

9612090327 18.

DEC 05 1996