

ENCLOSURE

NOTICE OF VIOLATION

Robert E. Lawrence-Berry, M.D.

License No. 47-15238-01

The following violation was identified during an inspection conducted on January 16, 1985. The Severity Levels were assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

License condition 15 requires that except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in the application and attachments dated September 22, 1982; and letter and attachments dated November 30, 1982. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

The letter dated November 30, 1982, states that the survey meter will be calibrated annually.

Contrary to the above, as of January 16, 1985, the survey meter had not been calibrated since January 21, 1983.

This is a Severity Level IV violation (Supplement VI).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: FEB 06 1985

8507230050 850206
REG2 LIC30
47-15238-01 PDR