

DELTA PENN CORPORATION
RD 6, BOX 494
LATROBE, PENNSYLVANIA 15650

MS 16
P2

June 10, 1985

Mr. Mike Varela
USNRC
Region I
631 Park Avenue
King of Prussia, PA 19406

Re: Application for license for Nuclear
Gauges and Belt Scales,
(Mail control #03671)

1. LOCATION: (A) The density gauge is to be located 17 feet above the main floor level on an 8" pipe. (B) The "raw coal feed" scale is to be located on the incoming conveyor on the top floor. (C) The "clean coal" scale is to be located on the clean coal conveyor on the ground floor, (D) The "refuse scale" is to be located on the refuse conveyor on the ground floor.

2. SIGNS: Radiation signs will be posted at each location with a danger warning.

3. BUILDING: The building is of steel construction on concrete piers with steel sheeting and insulation.

4. LOCK UP PROCEDURES: Lock up procedures will be implemented for "off-shift" or non-operating periods. Only three (3) people will be allowed to lock and unlock the shutters. (1) the Project Manager, Robert Walston, (2) the Lab Supervisor, Bruce Williams, (3) the Shift Control Room Foreman, Jack Washnock.

5. MAINTENANCE: All maintenance will be performed by the manufacturer personnell, Ohmart.

6. LEAK TEST: The Ohmart Leak Test Kit, model # LT will be used by the lab Supervisor, Bruce Williams, to complete the required leak test proceures. Mr. Williams will attend the Ohmart training school, July 9 and 10, 1985.

7. PLANT LOCATION: The plant is located in Unity Township, across the road and east from the Hostetter Post Office, Hostetter, Pa.

8507230407 850710
REG1 LIC30
37-20836-01 PDR


03671
"OFFICIAL RECORD COPY"

ML10

JUL 05 1985

8. EMERGENCY PROCEDURES: The operating personnell will have instructions to follow, in the event a scale or gauge is damaged (or fails)! (A) The area will immediately be roped off, (B) Signs posted to "Keep Away from the Area". (C) Call R. M. Walston and/or Bruce Williams.

Sincerely,

A handwritten signature in dark ink, appearing to read "Robert M. Walston", with a stylized flourish at the end.

Robert M. Walston
Project Manager

RMW/jav

DATE 4/22/85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME 11:17 ☒ A.M. ☐ P.M.

☒ INCOMING CALL

☒ OUTGOING CALL

☐ VISIT

PERSON CALLING

Bob Wallston

OFFICE/ADDRESS

Delta-PennCorp

PHONE NUMBER

EXTENSION

412-539-6667

PERSON CALLED

Mike Varela

OFFICE/ADDRESS

R-I

PHONE NUMBER

EXTENSION

CONVERSATION

SUBJECT

appl'n for license

030-21244

03120

SUMMARY

answer:

Location of use

Facility Descript.

Control of access

Lockout procedures

Leak testing procedures

+ Emergency Procedures

on 6/7/85, ~~Bob~~ On 6/7/85 Bob called back ~~at~~
(9:40 am) & I cleared up the above questions &
stressed documentation of training from Ohmart.

"OFFICIAL RECORD COPY"

ML10

REFERRED TO:

ACTION REQUESTED

☐ ADVISE ME OF ACTION TAKEN.

INITIALS

DATE

ACTION TAKEN

INITIALS

DATE