

# U. S. NUCLEAR REGULATORY COMMISSION MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter 1, Parts 30, 31, 32, 33, 34, 35, 36, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s); and to import such byproduct and source material. This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<b>Licensee</b>  1. Ted Forsi & Associates, Inc.  124 East Seventh Avenue 2. Anchorage, Alaska 99501		3. License number  50-19401-01
		4. Expiration date July 31, 1985
		5. Docket or Reference No.
6. Byproduct, source, and/or special nuclear material  A. Cesium 137  B. Americium 241	7. Chemical and/or physical form  A. Sealed Sources (Troxler Dwg. A-102112)  B. Sealed Sources (Troxler Dwg. A-102451)	8. Maximum amount that licensee may possess at any one time under this license  A. Not to exceed 10 millicuries per source  B. Not to exceed 50 millicuries per source
9. Authorized use  A. and B. For use in Troxler Model 3401, 3401B, 3411, and 3411B gauges to measure properties of materials.		

## CONDITIONS

10. Licensed material may be used at 124 East Seventh Avenue, Anchorage, Alaska and at temporary job sites of the licensee anywhere in the United States where the U. S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."

447677488

379

12319

U. S. NUCLEAR REGULATORY COMMISSION  
MATERIALS LICENSE  
Supplementary Sheet

Page 2 of 3 Pages

License Number 50-19401-01

Docket or  
Reference No. \_\_\_\_\_

CONDITIONS

12. Licensed material shall be used by, or under the supervision and in the physical presence of, Keith Mobley.
13. Sealed sources containing licensed material shall not be opened or removed from the portable moisture/density gauges by the licensee.
14. A. (1) Each sealed source containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source received from another person shall not be put into use until tested.  
(2) The periodic leak test required by this condition does not apply to sealed sources that are stored and not being used. The sources excepted from this test shall be tested for leakage prior to any use or transfer to another person unless they have been leak tested within six months prior to the date of use or transfer.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U. S. Nuclear Regulatory Commission, Region V, Office of Inspection and Enforcement, 1990 N. California Boulevard, Suite 202, Walnut Creek, California 94596, describing the equipment involved, the test results, and the corrective action taken.
- D. Tests for leakage and/or contamination shall be performed by the licensee or by other persons specifically authorized by the Commission or an Agreement State to perform such services.

U. S. NUCLEAR REGULATORY COMMISSION  
MATERIALS LICENSE  
Supplementary Sheet

Page 3 of 3 Pages

License Number 50-19401-01

Docket or  
Reference No. \_\_\_\_\_

CONDITIONS

15. The licensee shall conduct a physical inventory every six (6) months to account for all sealed sources received and possessed under the license. The records of the inventories shall be maintained for two (2) years from the date of the inventory for inspection by the Commission, and shall include the quantities and kinds of byproduct material, location of sealed sources, and the date of the inventory.
16. The licensee may transport licensed material or deliver licensed material to a carrier for transport in accordance with the provisions of Title 10, Code of Federal Regulations, Part 71, "Packaging of Radioactive Material for Transport and Transportation of Radioactive Material Under Certain Conditions."
17. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated June 16, 1980. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

JUL 18 1980

Date \_\_\_\_\_

For the U. S. Nuclear Regulatory Commission

Original Signed by  
by Material Licensing Branch

Division of Fuel Cycle and  
Material Safety  
Washington, D.C. 20555

U. S. NUCLEAR REGULATORY COMMISSION  
MATERIALS LICENSE  
Supplementary SheetPage 1 of 1 PagesLicense Number 50-19401-01Docket or  
Reference No. \_\_\_\_\_

Amendment No. 01

Ted Forsi & Associates, Inc.  
124 East Seventh Avenue  
Anchorage, Alaska 99501

In accordance with letter dated July 28, 1980, License Number 50-19401-01  
is amended as follows:

Conditions 12. and 17. are amended to read:

12. Licensed material shall be used by, or under the supervision and in the physical presence of, Keith Mobley or Nils Lindholm.
17. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated June 16, 1980 and letter dated July 28, 1980. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

AUG 27 1980

Date \_\_\_\_\_

For the U. S. Nuclear Regulatory Commission

Material Licensing Branch

by \_\_\_\_\_

Division of Fuel Cycle and  
Material Safety  
Washington, D. C. 20555

12319