

JUL 08 1985

Fettig Laboratories  
ATTN: Patricia J. Fettig, Ph.D.  
900 Godfrey, S.W.  
Grand Rapids, MI 49503

Dear Dr. Fettig:

Enclosed is Amendment No. 02 which renews NRC License Number 21-18943-01 in accordance with your request.

Per our telephone discussion on June 19, 1985, please note that this license authorizes storage only of the detector cell containing nickel-63. As long as the device is in storage you are excepted from the leak test requirements as specified in License Condition 13.E. Prior to transfer of the device, however, the detector must be tested for leakage.

If you desire to use the detector cell in the future for sample analyses, your license must be amended accordingly.

In view of these changes, please review your license carefully to assure that you understand and are in compliance with the terms and conditions contained therein. If you have questions, please contact our office by dialing (312) 790-5625.

Sincerely,

Original Signed By  
B. J. Holt  
Materials Licensing Section

Enclosure: Amendment No. 02

8507240345 850708  
REG3 LIC30  
21-18943-01 PDR

RIII

Holt/bs  
6/24/85