

NOV 24 1980

FCMLB:MAL
030-10233
(05652)

Divine Providence Hospital
ATTN: Lee B. Youngquist
Administrator
1004 Arch Street
Pittsburgh, PA 15212

Gentlemen:

This is in reference to your letter dated October 6, 1980, to amend your Material License No. 37-16031-01. In order to complete our review of your request, we need clarification on one item.

In order to amend your license to include Group III generators, you should verify that you will modify the linearity check of your dose calibrator to use an activity equivalent to that of the highest activity of use, e.g., first elution from a new generator when appropriate.

Upon receipt of this verification, we will be pleased to amend your license as requested. You should make your reply in duplicate, and refer to Control No. 05652.

Sincerely,

Michael A. Lamastra
Material Licensing Branch
Division of Fuel Cycle and
Material Safety

Std. Br. Dist.

ML18

"OFFICIAL RECORD COPY"

CRESS:WILL
MC#080469:gtw

FCMLB

MALamastra:gtw

11/14/80

11/17/80