



Veterans
Administration

June 7, 1985

MS 16
PS

Nuclear Regulatory Commission
Attn: Jenny Johansen
Region I
King of Prussia, PA 19406

SUBJ: Amendment to License No. 20-10184-01
Control No. 03-811

In reference to our telephone conversation on June 3, 1985 concerning the above subject, the following items should read as follows:

I. Section 10.2 Part A-3

"Instrument accuracy (at installation, annual intervals, and following a variation of greater than $\pm 5\%$ in instrument constancy check results)"

II. Section 17 - Nuclear Medicine Area Survey Procedures

1. "All elution, preparation, and injection areas will be surveyed daily, if radiopharmaceuticals are used that day, with an appropriately low-range survey meter and decontaminated if necessary. Records of these surveys will be kept."

Thank you.

CHOOMPOL MAHASAEN, M.D.
Chief, Laboratory Service/
Radiation Protection Officer

8507170667 850625
REG1 LIC30
20-10184-01 PDR

03811

"OFFICIAL RECORD COPY"

ML10

JUN 10 1985

6-3-85

TELEPHONE OR VERBAL CONVERSATION RECORD

TIME

11:00

☒ A.M.
☐ P.M.

☐ INCOMING CALL

☒ OUTGOING CALL

☐ VISIT

PERSON CALLING

Johansen

OFFICE/ADDRESS

USNRC Region I

PHONE NUMBER

EXTENSION

337-5215

PERSON CALLED

Dr. Chompot Mahasae

OFFICE/ADDRESS

RSO VA Med CTR Bedford

PHONE NUMBER

EXTENSION

8-840-0000

CONVERSATION

617 275-7500

SUBJECT

Amend. L/N 20-10184-01 CN 03811

SUMMARY

Deficiency: ① You must commit to an dose calibrator
Instrument accuracy test at least annually
- no commitment in application
② Area surveys - elution, injected, prep
areas - must be recorded.
- application states these need not be
recorded 10 CFR 20.401 requires survey records.

Dr Mahasae stated he would send
letter ~~with~~ correcting these 2 items of the
application.

ML10

"OFFICIAL RECORD COPY"

REFERRED TO:

CN 03811

ACTION REQUESTED

Letter from Licensee

☐ ADVISE ME OF
ACTION TAKEN.

INITIALS

DATE

6/3/85

ACTION TAKEN

INITIALS

DATE