

ENCLOSURE

NOTICE OF VIOLATION

Sussman Professional Associates of Virginia

License No. 45-23020-01

The following violation was identified during a videotape review conducted on October 31, 1984. The Severity Levels were assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

License Condition 15 of NRC License No. 45-23020-01 requires that the Lixiscope be used by the individuals specified in License Condition 12 on those portions of the human body upon which they are authorized by a State of Virginia license to practice podiatry medicine.

Contrary to the above, during a televised news interview conducted at your facility in late October 1984, a news reporter used the Lixiscope to demonstrate its function by performing an x-ray of his hand, thereby violating the requirements of License Condition 15.

This is a Severity Level IV violation (Supplement VI)..

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: NOV 23 1984

8507170674 841123
REG2 LIC30
45-23020-01 PDR