

MAY 23 1985

Docket No. : 030-20385
License No.: 50-23224-01
Control No.: 70172

Department of Health & Human Services
Arctic Investigations Laboratory
Center for Infectious Diseases
Centers for Disease Control
225 Eagle Street
Anchorage, Alaska 99501

Attention: Alan J. Parkinson, Ph.D.
Chief Immunologist

Gentlemen:

Enclosed is the NRC license which you requested.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify us so that we can provide appropriate corrections and answers.

Although your procedures now specify that laboratory areas must be surveyed by your personnel after each use of licensed material, your Radiation Safety Officer (RSO) should also conduct independent laboratory surveys at a specified frequency. The surveys should include both radiation level and contamination wipe tests. You should submit the procedure that your RSO uses within the next 30 days.

Any future correspondence relating to your license should specifically reference your license and docket numbers to expedite your inquiry.

Sincerely,

8507170694 850523
REGS LIC30
50-23224-01 PDR

Beth A. Riedlinger
Health Physicist (Licensing)
Nuclear Materials Safety Section

Enclosure: Attachment A
License No. 50-23224-01

OFFICE	RV	BAR					
SURNAME	Riedlinger	fr					
DATE	5/23/85						