

## MATERIALS LICENSE

Amendment No. 02

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		In accordance with application dated March 27, 1985	
1. Kelsey Hayes Company Fremont Foundry		3. License number 34-16091-01 is amended in its entirety to read as follows:	
2. 1017 Dickinson Fremont, OH 43420		4. Expiration date	July 31, 1990
		5. Docket or Reference No.	030-10354
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license	
A. Cesium-137	A. Sealed sources (3M Company Model No. 4F6S)	A. 2 sources not to exceed 750 millicuries each	
9. Authorized Use			
A. To be used in CE In-Val-Co Model B-20-50A source holder for level measurement.			

CONDITIONS

10. Licensed material shall be used only at the licensee's facilities located at 1017 Dickinson, Fremont, Ohio.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12. Licensed material shall be used by, or under the supervision of, Gary Morywas, Gene Hay, David McDiarmid, or Terry Gullett.
13. A. (1) Each sealed source containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source received from another person shall not be put into use until tested.
- (2) The periodic leak test required by this condition does not apply to sealed sources that are stored and not being used. The sources excepted from this test shall be tested for leakage prior to any use or transfer to another person unless they have been leak tested within six months prior to the date of use or transfer.

8508020169 850717  
REG3 LIC30  
34-16091-01 PDR

COPY

2  
ml38

DS 7/17/85

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License number 34-16091-01

Docket or Reference number  
030-10354

Amendment No. 02

- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U. S. Nuclear Regulatory Commission, Region III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137, describing the equipment involved, the test results, and the corrective action taken.
- D. The licensee is authorized to collect leak test samples in accordance with the procedures described in the licensee's application dated March 27, 1935 for analysis by Applied Health Physics, Inc. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.
14. Sealed sources containing licensed material shall not be opened or removed from their respective source holders by the licensee.
15. Installation, relocation, removal from service, maintenance, repair, and initial radiation survey of devices containing licensed material and installation, replacement, and disposal of sealed sources containing licensed material used in devices shall be performed only by the device manufacturer or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
16. The licensee shall operate each gauge within the manufacturer's specified environmental limits such that the shielding and shutter mechanism of the source holder is not compromised.
17. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated March 27, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

For the U.S. Nuclear Regulatory Commission

Date July 17, 1985

Original Signed  
By Patricia J. Whiston  
Materials Licensing Section, Region III

COPY