

Appendix

NOTICE OF VIOLATION

Lake County Health Department

License No. 13-16474-03

As a result of the inspection conducted on September 4, 1985, and in accordance with the General Policy and Procedures for NRC Enforcement Actions, (10 CFR Part 2, Appendix C), the following violations were identified:

1. License Condition 12 states that licensed material shall be used by or under the supervision of Andrew F. Livovich.

Contrary to this requirement, licensed material was used by an individual other than Andrew F. Livovich. Specifically, Andrew F. Livovich retired in June 1985, and licensed material was used by an individual not authorized by the license.

This is a Severity Level IV violation (Supplement VI).

2. License Condition 14a states that each chromatograph detector cell containing nickel-63 shall be tested for leakage and/or contamination at intervals not to exceed six (6) months.


Contrary to this requirement, the licensee failed to test each chromatograph detector cell containing nickel-63 for leakage and/or contamination every 6 months. Specifically, the last test for leakage and/or contamination was on February 15, 1985, and the licensee failed to leak test the detector cell again in August 1985.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

9/26/85

  
D. J. Sreniawski, Chief  
Nuclear Materials Safety Section 2

8510010467 850927  
REG3 LIC30  
13-16474-03 PDR