

30-14414



St. Vincent Medical Center

May 31, 1985

Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 61037

U.S. N.R.C.
LIC. FEE MGMT. BRANCH

85 JUN 26 PM 2:50

RECEIVED

Report of survey conducted in accordance with source deposition procedures required by the United States Nuclear Regulatory Commission.

License: St. Vincent Hospital and Medical Center
Radiology Department
2213 Cherry Street
Toledo, Ohio 43608

License Number: 34-01216-04

Expiration Date: November 30, 1988

Teletherapy Sealed Source: Source S-3945
2.0 cm diameter standard source
6828 Curies of Cobalt 60 on 5/3/83.

Survey Performed By: Kathryn J. Williford, M.S.
Medical Radiation Physicist
St. Vincent Medical Center
Certified, American Board of Radiology

Teletherapy Unit: Atomic Energy of Canada, Ltd.
Theratron 80

Survey Meter Used: Victoreen 471 Ion Chamber
Calibration Date, January, 1985
Source check performed; chamber response within acceptable limits.

8507190478 850628
REG3 LIC30
34-01216-04 PDR

Results of Area Survey: Background levels of radiation only found in teletherapy suite following source and machine removal.

Applicant *June 27*
Check No. *711 and*
Amount *no category*
Type of Fee *711 and*
Date Check Rec'd *6/26/85*
Received By *[Signature]*

tele survey
FEE EXEMPT term

Page Two

Date of Area Survey:

May 16, 1985

Documentation:

Service ^{record} ~~re-load~~ and copy of source disposal certificate from atomic energy of Canada, Ltd. is attached to this report.

Report Submitted By:

K. J. Williford, M.S.

K. J. Williford, M.S.
Medical Radiation Physicist
Certified, American Board of Radiology

CONTROL NO. 7 9 1 8 0



**Atomic Energy of Canada Limited
Commercial Products
SOURCE DISPOSAL CERTIFICATE**

TO WHOM IT MAY CONCERN:

This is to certify that the following source has been removed from the unit described herein, and returned to Atomic Energy of Canada Limited, Commercial Products, Ottawa, Ontario, Canada for disposal:

COBALT 60 OR CAESIUM 137 SEALED SOURCE	SERIAL NO. S-3495	DEPLETED URANIUM lb.	UNIT THERATRON 80.	UNIT SERIAL NO. 69
LOCATION OF UNIT				
ST. VINCENTS Hosp.				
2213 CHERRY ST.				
TOLEDO, OHIO. 43698				

Date: **May 15, 1985** Signed: **D. [Signature]**
A.E.C.L. Service Representative



Atomic Energy of Canada Limited

Medical Products

SERVICE RECORD

CUSTOMER ORDER NO. & DATE		COST CENTER		CHARGE POINT		PRODUCT ID	LOC.	SERVICE RECORD NO.	
90239		656					20	56544	B
REQUEST DATE	TIME	START DATE	TIME	COMPLETE DATE	TIME	MODEL & SERIAL NO.		S.B. HOURS	H.V. HOURS
		5-14-85	1200	5-16-85		T8D#69			
SERVICE REQUESTED						CUSTOMER			
REMOVE T8D#69 AND SHIP TO OTTAWA.						St Vincents Hosp Toledo Oh.			
SERVICE PERFORMED						LABOR HOURS			
REMOVED & SHIPPED UNIT.						STANDARD		OVERTIME	
						TRAVEL HOURS		STANDARD	
						TRAVEL EXPENSES		\$	
						TYPE OF SERVICE		CLINICAL DOWNTIME	
						BILLED SERVICE		SERVICE CONTRACT	
						SOURCE REPLACEMENT		SHOP REPAIR	
						UNIT INSTALL'N		REMANUFACTURE	
						WARRANTY		EXTRA ORDINARY	
FAILURE CODE						INVOICING INFORMATION			
						LABOR			
						TRAVEL			
						MATERIAL			
						FREIGHT			
						TAX			
						CUSTOMER TOTAL			
QUANTITY	U/I	PART NUMBER		DESCRIPTION		UNIT PRICE		THIS IS NOT AN INVOICE	
								INVOICING WILL FOLLOW IF REQUIRED AND WILL REFERENCE THIS SERVICE RECORD NUMBER	
								WARRANTY 30 DAYS PARTS AND LABOR	
ACCEPTED BY SIGNATURE				SERVICE REP. SIGNATURE				EMPLOYEE NO.	
Edward D. May				W. Hughes				0046	

CUSTOMER COPY