

SEP 13 1985

Therapy Services Incorporated
ATTN: Jerry L. Fogle
President
5409 Feagaville Lane
Frederick, Maryland 21701

Dear Mr. Fogle:

During a recent inspection conducted by members of my staff at the Frederick Cancer Research Center, we noted that your company had performed work on the source drive mechanism of a J. L. Shepherd Model 1 irradiator.

We have verified with the State of Maryland, Division of Radiation Control that your current Maryland State License only authorizes you to perform work on teletherapy units. Therefore, you must amend your Maryland State License to authorize service on specific irradiators in order to perform this work at Federal facilities and in states in which the Nuclear Regulatory Commission maintains jurisdiction of the use of byproduct material. In addition, you must file a properly filled out Form NRC-241 prior to conducting the work.

Thank you for your cooperation in this matter.

Sincerely,

Original Signed By:

John D. Kinneman

John D. Kinneman, Chief
Nuclear Material Safety
Section A

cc:
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
State of Maryland

bcc:
Region I Docket Room (w/concurrences)
Senior Operations Officer

JAC
RI:DRSS
Rowe/mjh

9/6/85

JK
RI:DRSS
Kinneman

9/7/85

OFFICIAL RECORD COPY

DL THERAPY SERVICES - 0001.0.0
09/04/85

8510020003 850913
REG1 LIC30

PDR

RETURN ORIGINAL TO
REGION I

10
IE:07