



AT&T
Technology Systems

C. R. Meetsma
General Manager
Allentown Works

555 Union Boulevard
Allentown, PA 18103
Phone (215) 439-6211

July 2, 1985

MR. THOMAS T. MARTIN, Director
Division of Radiation Safety & Safeguards
U.S. Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, PA 19406

Dear Sir:

On 6/26/85, we received a letter from you concerning certain activities which were not being conducted in full compliance with NRC requirements. Our disposition is as follows:

Item A - Security

The door to the tracer-flo activation room has been equipped with a positive lock. Likewise, the cabinet containing failed devices has also been fitted with a lock.

Our security requirements have been reviewed with all employees in the area.

Item B - Calibration

Our survey meters have been calibrated by AT&T-Bell Laboratories of Murray Hill, NJ, with Cs-137 and 40 kV_p x-rays. Arrangements have been made to perform this calibration annually.

We trust you are in agreement with these actions. Thank you for your interest in our radiation protection program.

Very truly yours,

8510020093 850927
REG1 LIC30
37-03100-02 PDR