

ENCLOSURE

NOTICE OF VIOLATION

Advex Corporation

License No. 45-16452-01

The following violations were identified during an inspection conducted on March 27-28, 1985. The Severity Levels were assigned in accordance with the NRC Enforcement Policy (10 CFR Part 2, Appendix C).

1. 10 CFR 34.24 requires a licensee to calibrate each radiation survey meter, used to conduct physical radiation surveys, at intervals not to exceed three months.

Contrary to the above, the licensee did not calibrate radiation survey meter No. 4938, a meter used to conduct physical radiation surveys, between August 16, 1984 and February 8, 1985, an interval in excess of three months.

This is a Severity Level IV violation (Supplement VI).

2. 10 CFR 34.33(a) forbids a licensee to permit any individual to act as a radiographer unless, at all times during radiographic operations, each such individual wears a direct reading pocket dosimeter.

Contrary to the above, on March 27, 1985, the licensee permitted a radiographer, who conducted radiographic operations at the Langley Research Center, to perform operations without having a pocket dosimeter in his possession.

This is a Severity Level IV violation (Supplement VI).

Pursuant to 10 CFR 2.201, you are required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violations; (2) the reasons for the violations if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved.

Security or safeguards information should be submitted as an enclosure to facilitate withholding it from public disclosure as required by 10 CFR 2.790(d) or 10 CFR 73.21.

Date: APR 08 1985

8507190580 850408
REG2 LIC30
45-16452-01 PDR