

APPENDIX A

NOTICE OF VIOLATION

Camden County Municipal Utilities Authority
Camden, New Jersey 08101

Docket No. 9990001
General License

As a result of the inspection conducted on June 26, 1985, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

10 CFR 31.5(c)(2) requires that any device possessed pursuant to 10 CFR 31.5(a) be tested for leakage of radioactive material at no longer than three year intervals.

Contrary to the above, as of June 26, 1985, sealed sources containing 500 millicuries each of cesium-137 had not been tested for leakage since April 1982, an interval of more than three years.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, Camden County Municipal Utilities Authority is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

8510070074 850930
IE GA999 EUTCAMMU
99990001 PDR

OFFICIAL RECORD COPY

DL CAMDEN - 0003.0.0
09/20/85