

JUL 25 1985

30-09632

SDS Biotech Corporation
ATTN: Mr. Richard J. Leisman
Radiological Safety
and Hazardous Waste Officer
7528 Auburn Road
Box 348
Painsville, OH 44077

License No. 34-000583-04

Gentlemen:

Thank you for your inquiry dated July 15, 1985. We have reviewed your license and its supporting documents and can find no restrictions on replacement of detector cells. However, Condition 15, specifically prohibits you from opening these detector cells for foil replacement or any other purpose. Should you desire to open detector cells for foil replacement, you should submit an amendment request along with the appropriate fee and include a copy of your step-by-step procedures for these activities as well as what precautions you will take to minimize extremity exposure.

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

Sincerely,

Original Signed By
William J. Adam, Ph.D.
Materials Licensing Section

RIII
WJA
Adam/cm
07/23/85

8508070177 850725
REG3 LIC30
34-00583-04 PDR

11 ml 31
AS 7/25/85