

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number
12-10142-01

Docket or Reference number
030-04189

Amendment No. 07

Phoenix Chemical Laboratory, Inc.
3953 West Shakespeare Avenue
Chicago, IL 60647

In accordance with letter dated June 14, 1985, License Number 12-10142-01 is amended as follows:

Condition 13.A. is amended to read:

13. A. Each chromatograph detector cell containing Nickel-63 shall be tested for leakage and/or contamination at intervals not to exceed six months; except those detector cells as specified by the manufacturer and specifically authorized by the Commission or an Agreement State may be leak tested at intervals not to exceed three years. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a detector received from another person shall not be put into use until tested.



8507220272 850628
REG LIC30
12-10142-01 PDR

For the U.S. Nuclear Regulatory Commission

Date June 28, 1985

Original Signed
By Patricia J. Whiston
Materials Licensing Section, Region III

COPY

2
me30

BS 7/9/85